

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
<i>Hewitt Fertility Centre</i>	
Owner: K Schnauffer	Author: S Brooks

## Ormskirk District General Hospital



# DIAGNOSTIC SEMINOLOGY SERVICES

# USER MANUAL

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

## CONTENTS

<b>CONTENTS</b> .....	2
1. INTRODUCTION.....	3
2. CONTACT DETAILS.....	3
3. LOCATION .....	4
4. OPENING TIMES.....	4
5. HOW TO REQUEST A SEMEN ANALYSIS .....	4
5.1 Referrals from outside Liverpool Women's Hospital.....	4
6. INSTRUCTIONS FOR PRODUCTION OF SEMEN SAMPLES .....	5
7. INSTRUCTIONS FOR TRANSPORTATION OF SEMEN SAMPLES .....	5
8. ROUTINE TESTS PROVIDED .....	5
10. EXAMINATION OF POST-VASECTOMY SAMPLES.....	6
11. INTERPRETIVE COMMENTS & TERMINOLOGY .....	7
12. REPORTING OF RESULTS.....	7
13. TURNAROUND TIME .....	7
14. PROVISION OF CLINICAL ADVICE/Complaints.....	7
15. Measurement of Uncertainty .....	8
16. Procedure 'flow-diagram' with approximate timescales .....	8
17. REFERRAL PROFORMA.....	10
18. SAMPLE OF LETTER SENT TO PATIENT.....	11
19. SEMEN ANALYSIS FORM.....	12
20. PATIENT INSTRUCTION LEAFLET.....	13
21. OFF-SITE PRODUCTION FORM.....	14-15

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

## 1. INTRODUCTION

This manual describes the policies, procedures and repertoire of the Seminology Laboratory, Antenatal and Gynaecology Department at Ormskirk. This department is manned part-time and is therefore under the direction and facilitated administratively by the Andrology laboratory, Hewitt Fertility Centre at the Liverpool Women's Hospital. We work towards the standards outlined by WHO 2010 Examination and processing of human semen and ISO 15189:2012 Medical laboratories – Requirements for quality and competence.

## 2. CONTACT DETAILS

The postal address of the Andrology Laboratory is

Andrology Laboratory  
Hewitt Centre for Reproductive Medicine  
Liverpool Women's Hospital  
Crown Street  
Liverpool  
L8 7SS  
Tel: 0151 702 4214  
Fax: 0151 702 4152

The Andrology Laboratory is under the direction and management of Dr Stephen Troup (Scientific Director) (ISO 15189:2012 4.1.14) and Ms Karen Schnauffer (Consultant Embryologist), respectively. The Andrology Laboratory is supervised by Mrs Stephanie Brooks (Andrology Supervisor) and staffed by Miss Laurie Wishman and Mr Andrew Allan (Andrology practitioners) trained Andrologists, Mrs Judith Cunningham, Mr Paul Mallanaphy and Ms Amanda Hall supported by Mrs Cheryl Thomas (Medical Laboratory Assistant) & Miss Jennifer Edge (Scientific Andrology Assistant).

We are under the supervision in regards to Quality management by Mrs Sharon Fensome-Rimmer who directs the service with regards to maintenance of our Quality management system. (ISO 15189:2012 4.12.7, 4.2)

This, in conjunction with the team above includes meeting the needs of our users (ISO 15189:2012 4.1.2.2, 4.4, 4.14.3), Service level agreements & third party agreements ((ISO 15189:2012 4.4), complaints (ISO 15189:2012 4.8), assessment of user feedback (ISO 15189:2012 4.14.3), review input (ISO 15189:2012 4.15), training of staff and competency assessment (ISO 15189:2012 5.1.5, 5.1.6), maintenance of facilities(ISO 15189:2012 5.2),Equipment maintenance (ISO 15189:2012 5.3), reagents and consumables (ISO 15189:2012 5.3.2), pre-examination processes (ISO 15189:2012 5.4), examination processes (ISO 15189:2012 5.5), ensuring quality of examination of results (ISO 15189:2012 5.6), post examination processes, (ISO 15189:2012 5.7), reporting of results (ISO 15189:2012 5.8). This list is not exhaustive and may include additional items.

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

### 3. LOCATION

(ISO 15189:2012 5.4.2 A)

The Seminology Laboratory is situated on the ground floor of Ormskirk District General Hospital. From the main entrance of the hospital proceed through the Ante-natal Clinic entrance and turn first right, the laboratory is located on your left.

### 4. OPENING TIMES

(ISO 15189:2012 5.4.2 C)

The Seminology Laboratory at Ormskirk is open 0830-1630 Monday (although usually closed on Bank Holidays). At the Liverpool Women's Hospital there is a telephone service and answer-machine in operation when the laboratory is closed, on 0151-702-4214. This service can be used regarding appointments which need to be cancelled or changed. There is no out-of-hours service, unless exceptional circumstances prevail.

### 5. HOW TO REQUEST A SEMEN ANALYSIS

(ISO 15189:2012 5.4.3)

#### 5.1 Referrals from outside Liverpool Women's Hospital

Clinicians wishing to arrange an appointment for a semen analysis should complete a 'Semen Analysis Request Form' (See Appendix 1). It is essential that this form is completed in full and is legible.

The 'Semen Analysis Request Form' should be faxed to 0151 702 4152.

On receipt of this form the Andrology Laboratory will arrange an appointment by sending the patient a pack containing the following:

- i. covering letter (see Appendix 2)
- ii labelled semen collection pot(s)
- iii. labelled 'Semen Analysis' form (Appendix 3 & 4) including a date and time of their appointment
- iv. labelled 'Record of Semen Production' form (Appendix 5)
- v. Patient Information Leaflet (Appendix 6)

When the patient makes contact with the Andrology Laboratory a convenient appointment should be made.

NB. Non-routine appointments must be arranged by contacting the Andrology Laboratory.

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

## 6. INSTRUCTIONS FOR PRODUCTION OF SEMEN SAMPLES

### (ISO 15189:2012 5.4.4.2)

Instructions for the production of semen samples are detailed in leaflet entitled 'Instructions for the production of semen samples' [see Appendix 2]. The date of the appointment should also be entered on this form.

Patients should be advised to follow the instructions in this leaflet in order to optimise the semen sample that they produce.

Patients should be provided with a suitable sample collection vessel and plastic transportation bag.

**Patients should be advised to only use the collection vessel provided.**

The following instructions are contained within this leaflet:

The patient should

- i. not ejaculate for between 2 and 7 days prior to their appointment
- ii. only use the pot provided to collect their sample
- iii. clearly label the pot with their name, date and time of sample production and the number of days since they last ejaculated
- iv. empty their bladder before producing the semen sample
- v. produce the sample by masturbation and not by 'withdrawing' after intercourse
- vi. not use a condom to collect the sample as condoms can adversely affect sperm\*
- vii. attempt to collect ALL of the sample into the pot and advise a member of the Seminology staff if any of the sample is not collected
- viii. \* If the patient is unable to produce a semen sample by masturbation, then special condoms (a 'Male Factor Pack') are available by prior arrangement with the Seminology Laboratory.

## 7. INSTRUCTIONS FOR TRANSPORTATION OF SEMEN SAMPLES

### (ISO 15189:2012 5.4.2 H)

If the patient is producing the sample 'off-site' he should be instructed to not expose the sample to extremes of temperature, by carrying it in an inside pocket if possible.

The patient should be instructed to deliver the sample to the Seminology Laboratory **WITHIN ONE HOUR** of production.

## 8. ROUTINE TESTS PROVIDED

### (ISO 15189:2012 5.4.2 D)

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

The Seminology Laboratory provides a range of diagnostic seminology tests and follows recommendations made by the World Health Organisation (WHO Laboratory manual for the examination and processing of human semen, Fifth Edition, 2010), the British Andrology Society, Association of Biomedical Andrologists and the Association of Clinical Embryologists.

A routine semen analysis will assess the following seminal parameters:

SEMINAL PARAMETER	COMMENTS
<b>Liquefaction</b>	A qualitative assessment of how liquefied the ejaculate has become. Measured at least 30 minutes post-ejaculation.
<b>pH</b>	The pH of the ejaculate. Measured at least 30 minutes post-ejaculation.
<b>Appearance</b>	A qualitative assessment of the visual appearance of the ejaculate. Eg Normal, opaque etc
<b>Presence of round cells</b>	A quantitative assessment of the number of non-sperm cells in the ejaculate (NB no differentiation is made between non-sperm round cells and leucocytes). Reported as millions round cells per ml of ejaculate
<b>Presence of acellular debris</b>	A qualitative assessment of the amount of acellular debris present in the ejaculate. Reported as 0, +, ++ or +++
<b>Ejaculate volume</b>	The volume of the ejaculate measured in millilitres (ml)
<b>Sperm concentration</b>	Millions sperm per ml of ejaculate (millions/ml)
<b>Sperm morphology</b>	Percentage of sperm with 'normal' morphology (%)
<b>Sperm motility</b>	The motility of at least 200 sperm is assessed (at 37°C) and expressed as the percentage showing progressive, non-progressive or immotile
<b>Presence of agglutination</b>	A qualitative assessment of the numbers of sperm 'sticking' to each other. Reported as 0, +, ++ or +++.
<b>Sperm viability</b>	Percentage of viable sperm (only measured if sperm motility <90%)

## 10. EXAMINATION OF POST-VASECTOMY SAMPLES

The guidelines issued by the British Andrology Society for the examination of post-vasectomy semen samples are followed throughout (P Hancock & E McLaughlin for the British Andrology Society, 2002, J.Clin.Path., p812-816)

The Seminology Laboratory will report any observations including the presence of very low numbers of immotile sperm. If greater than 50 non-motile sperm are seen on 10ul, a concentration will be performed to assist clinicians in giving 'special clearance'

The Seminology Laboratory will report any observations including the presence of very low numbers of immotile sperm.

Document Code: QMS-AND-UM-6	
Document Title: Semenology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

It is left to the clinical judgement of the referring clinician to deem patients 'fertile' or 'infertile' on the basis of semen analysis results, although clinical advice will gladly be provided on request (see 'Provision of Clinical Advice' below).

## 11. INTERPRETIVE COMMENTS & TERMINOLOGY

Term	Definition
<b>Aspermia</b>	No sample produced on ejaculation
<b>Azoospermia</b>	No sperm present in ejaculate
<b>Cryptozoospermia</b>	No sperm observed on initial examination <b>but</b> very low numbers observed following centrifugation concentration and examination of entire ejaculate
<b>Oligozoospermia</b>	<15 million sperm per ml of ejaculate
<b>Asthenozoospermia</b>	<32% grade A motility <b>or</b> <40% grade A + grade B motility
<b>Teratozoospermia</b>	<4% normal forms
<b>Haemospermia</b>	Presence of blood in the ejaculate
<b>Incomplete sample collection</b>	Patient has failed to collect entire ejaculate

In addition, other self-explanatory interpretative comments may be added.

## 12. REPORTING OF RESULTS

(ISO 15189:2012 5.8)

A 'Semen Analysis Report Form' is generated by the Hewitt Centre 'IDEAS' database and returned, by post, to the referring clinician.

### VERBAL RESULTS WILL NOT BE GIVEN OUT UNDER ANY CIRCUMSTANCES

The Semenology Laboratory endeavours to return results within 2 weeks of patient attendance.

## 13. TURNAROUND TIME

The Andrology Laboratory endeavours to return results within 2 weeks of patient attendance.

## 14. PROVISION OF CLINICAL ADVICE/COMPLAINTS

Clinical advice on any aspect of the diagnostic (or therapeutic) services provided by the Andrology Laboratory can be obtained from

Dr Stephen Troup, Scientific Director

0151 702 4173

Mr Richard Russell, Consultant Clinical Andrologist

0151 702 4215



Document Code: QMS-AND-UM-6	
Document Title: Semenology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

Andrology Laboratory

0151 702 4214

Or by e-mail enquiry to [stephen.troup@lwh.nhs.uk](mailto:stephen.troup@lwh.nhs.uk)/ [Karen.schnauffer@lwh.nhs.uk](mailto:Karen.schnauffer@lwh.nhs.uk)


Complaints should be directed to the Quality Manager or Scientific Director at the Hewitt Fertility Centre, Liverpool Womens Hospital, Crown Street, Liverpool L8 7SS. (ISO 15189:2012 4.8),

## 15. MEASUREMENT OF UNCERTAINTY

(ISO 15189:2012 5.5.1.3, 5.5.1.4, 5.5.3)

Clinicians and scientists are generally comfortable with the concept of uncertainty in relation to a blood test to determine for example a hormone level, but of course, a semen analysis comprises a combination of different test results. As such it is important to consider the measurement of uncertainty in relation to semen analysis testing and the mechanisms that are in place to attempt to minimise uncertainty of measurement when assessing semen samples. Therefore we have produced a document SCI-POL-1 Measurement of uncertainty in Semen analysis that we ask that you read. It includes a section at the back with bullet points that you are asked that you consider when interpreting the results that we provide.

## 16. PROCEDURE 'FLOW-DIAGRAM' WITH APPROXIMATE TIMESCALES





Document Code: QMS-AND-UM-6	
Document Title: Semiology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

Attach Patient Label: Male name: Date of Birth: Mobile Tel. no. Address:	Attach Patient Label: Female name: Date of Birth: Mobile Tel. No. Address:
NHS No.	NHS No.

**Details of referring Clinician** Please use Practice Stamp

Clinicians Name:

Practice address:

**Please tick appropriate boxes**

GOPD     Hewitt Centre     GP     Aintree     Ormskirk

This referral is    NHS     PP     Date of appointment if known \_\_\_\_\_

Reason for referral:    Fertility     Sperm Freeze

(please tick)    Vasectomy     Vasectomy reversal

Date performed: \_\_\_\_\_    Date performed: \_\_\_\_\_

**Is this patient High risk (i.e. Virus positive)**

Yes (please give details) \_\_\_\_\_  
 No

**Additional information (i.e. History of violence/aggression)** \_\_\_\_\_

**Please post or fax this request form to:**

Andrology Laboratory  
 Hewitt Fertility Centre  
 Liverpool Women's Hospital  
 Crown Street  
 Liverpool  
 L8 7SS  
 Tel No. 0151-702-4214  
 Fax No. 0151-702-4152

**PLEASE FILL IN ALL DETAILS CORRECTLY AND LEGIBLY AND YOUR PATIENT WILL BE CONTACTED TO ARRANGE THEIR APPOINTMENT. INCOMPLETE OR ILLEGIBLE FORMS WILL NOT BE PROCESSED.**

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

Dear Sir

**Re. Semen Analysis**

Your doctor has asked us to contact you in order to arrange an appointment to attend for a semen analysis test.

Please find enclosed a pack containing your sample pot(s) and paperwork. On one of these forms it indicates the date and time of your appointment.

Please contact us on **0151 702 4214** to confirm this appointment or rearrange if not convenient. If you **fail** to confirm this appointment, you risk losing your appointment slot. If telephoning us to confirm your appointment please feel free to leave a message on the answer machine. We may contact you by telephone, if you do not want us to do that, please inform us of that when you call. Please note that all appointments are Monday, 9.00am–9.30am or 12 noon-12.30pm.

Please also find enclosed some instructions. It is important that you read these instructions and complete the paperwork. Your results will be available from your doctor within 7-10 days from your Consultant at your next clinic appointment.

**Please note that your appointment is for the Seminology Laboratory, Women and Children’s Centre, Ormskirk Hospital, Wigan Road, Ormskirk, L39 2AZ.**

With kindest regards.  
Yours sincerely

Dr Stephen Troup  
Scientific Director

Document Code: QMS-AND-UM-6	
Document Title: Semiology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

Date of analysis:..... Lab code no.:.....

Male name .....DOB:.....Hospital no. ....

Female name .....DOB:.....Hospital no.....

Please circle appropriate clinic:

Ormskirk      HFC      GP      Vas      Referring clinician:..... NHS / PP

Sample pot Lot No.....	
Time of sample production .....	Analysis interval .....mins
Abstinence period.....(days) Appearance .....Viscosity : Viscous / Non-viscous	
pH .....	Round Cells ..... x 10 <sup>6</sup> /ml
<p>Volume: ..... ml          _____ - _____ -0.3= _____          Weight-Weight before-0.3</p> <p>Concentration:..... x 10<sup>6</sup>/ml (Performed by: )</p> <p>Total sperm number ..... M/ejaculate</p> <p>Morphology: ..... % (Performed by: )</p>	<p>Motility at .....°C (Performed by</p> <p>Grade A.....% (progressive)</p> <p>Grade B.....% (progressive)</p> <p>Grade C ..... % (non-progressive)</p> <p>Grade D ..... % (non-motile)</p> <p>Average speed.....M/sec</p>
<p>Agglutination      1      2      3      4</p> <p>Vitality ..... % (Performed by: )</p>	

Comments	Analysed	Analyst	Date
	IDEAS		
	2 <sup>nd</sup> person verified		
	Returned		

**Reference ranges WHO 2010**

Volume	1.5 mls or more
pH	7.2 or more
Concentration	15 million sperm per ml or more
Total sperm number	39 Million sperm per ejaculate
Total motility (A, B & C)	40 % (38-42%)
Progressive motility (A&B)	32 % (31-34%)
Sperm Morphology	4% or more
Vitality	58 % (55-63%) only measured when 90% non-motile

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

### Instructions for the production of semen samples

Your appointment is at the Seminology Laboratory on the ground floor of Ormskirk Hospital at the following date and time -

Date
------

Time
------

**Please read the following instructions carefully before producing your semen sample**

- Do not ejaculate for **2 – 7** days before your appointment.
- Produce your sample by masturbating into the pot provided.
- Samples can be produced at home or at the facilities available in the Seminology Laboratory. If producing at home, the sample must be delivered to the laboratory within one hour and at the date and time given above.
- Please complete the '**Record of Sperm Production Form**'.
- Results will be given by the referring clinician at your follow up appointment.
- Do not produce your sample using a condom, lubricant, the withdrawal method or by any other means other than masturbation.
- Do not expose to extremes of temperature.
- If you are unable to ejaculate by masturbation, then special condoms designed specifically for the collection of semen samples are available, please ask.

**If you do not attend your appointment there may be a 6-8 week wait for a further appointment which may delay your attendance at clinic.**

If you are unsure about any of the above points, or are unable to attend this appointment then please contact the Andrology laboratory on  
**0151 - 702 – 4214**

Document Code: QMS-AND-UM-6	
Document Title: Seminology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

**SECTION I – TO BE COMPLETED BY MALE OR FEMALE PARTNER**

I confirm that the sample container is correctly labelled with my/my partners correct details including name, D.O.B. and address (if applicable).

Signature of patient/ partner confirming details on sample pot \_\_\_\_\_

Date \_\_\_\_\_

**Section 2 - Please complete this section if you are the man who has produced the sample**

Your name \_\_\_\_\_ Date of birth \_\_\_\_\_

Partner's name \_\_\_\_\_ Date of birth \_\_\_\_\_

Your address \_\_\_\_\_

Where was your sample produced *At home / At the Hewitt Fertility Centre* (Please delete as appropriate)

If at home, what time was your sample produced \_\_\_\_\_

How many days is it since you last ejaculated \_\_\_\_\_

Was any of the sample spilled during collection? Yes\* / No

Have you been ill during last 3 months? \_\_\_\_\_ (eg Flu)

**Do you have or have you ever been told that you have HIV, Hepatits B or Hepatitis C Yes/ No**

I confirm the following with regard to the semen sample that I have handed to staff at the Hewitt Fertility centre today:

- i. That the sample was produced by me
- ii. That the sample has not been tampered with since its production
- iii. That the sample was produced at the time specified above
- iv. I am happy for any surplus specimen to be used for teaching and/or quality assurance purposes

Signature of patient \_\_\_\_\_ Date \_\_\_\_\_

**\* If some of the sample was lost during collection please inform a member of the laboratory staff**

Document Code: QMS-AND-UM-6	
Document Title: Semiology Diagnostic User Manual Ormskirk District General Hospital	
Version No: 3	
Date of issue: 16.07.2015	Date of review: 16.07.2016
Hewitt Fertility Centre	
Owner: K Schnauffer	Author: S Brooks

**Section 3 – Please complete this section if you are delivering the sample on behalf of your partner**

Your name \_\_\_\_\_ Date of birth \_\_\_\_\_

Your partner's name \_\_\_\_\_ Date of birth \_\_\_\_\_

Your partners address \_\_\_\_\_

What time did your partner produce the sample? \_\_\_\_\_

How many days is it since your partner last ejaculated \_\_\_\_\_

Did your partner spill any of the sample during collection? Yes\* / No

Has your partner been ill during last 3 months? \_\_\_\_\_ (please describe e.g. flu)

**Does your partner have or have they ever been told that they have HIV, Hepatitis B or Hepatitis C Yes/ No**

Please confirm the following with regard to the semen sample that you have handed to staff at the Hewitt Fertility centre today:

- v. The sample was produced by my partner named above
- vi. The sample has not been tampered with since its production
- vii. That the sample was produced at the time specified above

Signature of person delivering sample \_\_\_\_\_ Date \_\_\_\_\_

**\* If some of the sample was lost during collection please inform a member of the laboratory staff**

**SECTION 4 – TO BE COMPLETED BY STAFF RECEIVING SAMPLE**

I confirm that the paperwork belonging to the patient named above was handed to me at the time specified below, and the patient's details have been verbally confirmed.

Signature of staff member receiving paperwork \_\_\_\_\_

Time sample received \_\_\_\_\_ Date: \_\_\_\_\_

I confirm that I have received the sample from the patient named above and that the sample was appropriately labelled.

Signature of staff member receiving sample \_\_\_\_\_

Time sample received \_\_\_\_\_ Date: \_\_\_\_\_